



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

HELIUM LEAK TESTING, INC. CALIBRATION LABORATORY  
3910 Royal Ave, Unit A  
Simi Valley, CA 93063  
Jeff Markel Phone: 818 349 5690

CALIBRATION

Valid To: March 31, 2021

Certificate Number: 4217.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,4</sup>:

I. Fluid Quantities

| Parameter              | Range  | CMC <sup>2,3</sup> (±) | Comments   |
|------------------------|--|------------------------|--|
| Gas Flow Rate – Helium | [(1 x 10 <sup>-10</sup> ) to (1 x 10 <sup>-3</sup> )] Atm cc/sec | 4.0 %                  | Mass spectrometer comparison calibration systems |

II. Mechanical

| Parameter                  | Range   | CMC <sup>2,3</sup> (±) | Comments   |
|----------------------------|---|------------------------|--|
| Vacuum Gauges – All Gauges | [(1 x 10 <sup>-3</sup> ) to Atm] Torr                     | 4.4 %                  | Vacuum gauge measurement system w/ capacitance manometers static comparisons |
|                            | [(1 x 10 <sup>-7</sup> ) to (1 x 10 <sup>-3</sup> )] Torr | 4.4 %                  | Vacuum gauge measurement system w/ spinning rotor gauge dynamic comparison   |

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC, the percent is defined as the percentage of reading.

<sup>4</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



## *Accredited Laboratory*

A2LA has accredited

# HELIUM LEAK TESTING, INC. CALIBRATION LABORATORY

*Simi Valley, CA*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 22<sup>nd</sup> day of April 2019.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 4217.01  
Valid to March 31, 2021

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*